

Declaration of Conformity

**Declaration of Conformity in accordance with the Radio and
Telecommunications Terminal Equipment Act (FTEG) and Directive
1999/5/EC (R&TTE Directive)**

The manufacturer / responsible person: Baros GmbH • Kleine Düwelstr. 21 • 30171 Hannover declares that the product a-rival Oximeter oQsy comply with the essential requirements of §3 and the other relevant provisions of the FTEG (Article 3 of the R&TTE Directive), when used for its intended purpose.

Health and safety requirements pursuant to § 3 (1) 1. (Article 3(1) a)).

Harmonised standards applied EN 60950-1:2005 (2nd Edition) +
A1:2009 and EN 60950-1:2006 + A11:2009 + A1:2010 + A12:2011, EN
62479:2010

Protection requirements concerning electromagnetic compatibility § 3(1)(2), (Article 3(1)(b))
Harmonised standards applied ETSI EN 301 489-1 V1.9.2:2011
ETSI EN 301 489-3 V1.6.1:2013, EN 55024:2010 (IEC61000-4-2: 2008, IEC61000-4-3: 2010,
IEC61000-4-8: 2009), EN 55022: 2010,

Measures for the efficient use of the radio frequency spectrum.
Air interface of the radio systems pursuant to § 3(2) (Article 3(2))
Harmonised standards applied ETSI EN 300 440-1 V 1.6.1:2010,
ETSI EN 300 440-2 V 1.4.1:2010

Address: Baros GmbH, Kleine Düwelstr. 21, D-30171 Hannover

Phone, fax, e-mail +49 (0)511 / 988 71 85, Fax: +49 (0)511 / 988 74 58,
info@baros-gmbh.de



(Alexander Osberger, Managing Director)

Hannover, February 16. 2014

Place & date of issue